

INTOX EC/IR II
Quality Assurance Worksheet

ins
nsk

Instrument Serial Number 010599 Worksheet Start Date 6/10/2015
Location Prince William County Adult Detention Center
Address 9320 Lee Ave., Manassas, VA 20110
DFS Technician Heather Stanton License No. 18910
☐ Laboratory ☒ On-Site

Site Specification: No detrimental environmental conditions exist. ☒

Instrument Barometer (mm HG) 749 Reference Barometer (mm HG) 753
Reference Barometer(RB)Serial # 009113 RB Calibration Due 4/16/2016

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.300		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.100		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	


Dry gas standard Lot No. (with tank no.) AG400202-32

☐ Replaced dry gas standard (+O-ring)

☐ Installed at Location

☐ Removed to DFS-Central

Supplies
Mouthpieces
Certificates of Analysis
Operator Worksheet
Other:



Notes:

Instrument Serial Number

010599

Certification Date

☐ Calibrated☐ Certified☐ Measurement Assurance Check☒ Instrument Test☐ Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Heath W. Smith

Date

6/10/15

Issuing Analyst

Melissa Kennedy

Date

6/11/2015




COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

hmo
NJK

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
--	--------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY STANTON, HEATHER, M		AGENCY DFS Central Lab
DFS LICENSE NUMBER 18910	LICENSE EXPIRES 10/01/2016	DATE TEST CONDUCTED 06/10/2015
TEST EQUIPMENT NUMBER 010599		

RESULTS: TIME SAMPLE TAKEN 11:01 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20 _____.

BREATH TEST OPERATOR

☐ I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____ SUBJECT'S SIGNATURE

☐ SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____ OPERATOR'S SIGNATURE



IntoxNet MIS Report

Report Generated 10 Jun 2015 at 14:53

Handwritten initials: HNS and WSK

Test Results

Instrument Serial Number 010599

Test # 000629 Subject Test

Test Location 1 Department of
Test Date 10 Jun 2015

Test Location 2 Forensic Science
Test Time 10:55
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name STANTON
Agency DFS Central Lab

Operator's First Name HEATHER

Operator's Middle Initial N

Card Serial Number 118910

License Number 18910

Effective Date 10/01/2014

Expiration Date 10/01/2016

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 10 Jun 2015 End Time 11:03

Result Time 11:01

Result Date 10 Jun 2015

Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 10:55

Data Type BLK

Sample Value 0.000

Sample Time 10:56

Data Type CHK

Sample Value 0.098

Sample Time 10:57

Data Type BLK

Sample Value 0.000

Sample Time 10:58

Data Type SUBJ

Sample Value 0.000

Sample Time 10:59

Data Type BLK

Sample Value 0.000

Sample Time 11:00

Data Type SUBJ

Sample Value 0.000

Sample Time 11:01

Data Type BLK

Sample Value 0.000

Sample Time 11:03

Standard Type Dry Gas Std

Standard Value 0.099

Standard Lot Number AG400202-32

Standard Expiration Date 01/02/2016

Tank Pressure 566

Barometric Pressure 749 mmHg

Blow Sample Number 1 Blow Duration 3.16 sec

Blow Volume 1754 cc End-of-Blow Time 10:59

Blow Sample Number 2 Blow Duration 3.31 sec

Blow Volume 1770 cc End-of-Blow Time 11:01

Tamper Evident Stamp 6b75126b

Test Status Code 0

Test Status Success